UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/644,777	08/23/2000	Roger P. Jackson	10,033	1641
7590 12/26/2007 John C McMahon P O Box 30069			EXAMINER STOKES, CANDICE CAPRI	
	3732			
			MAIL DATE	DELIVERY MODE
			12/26/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

	Application No.	Applicant(s)					
, and the second	09/644,777	JACKSON, ROGER P.					
Office Action Summary	Examiner	Art Unit					
	Candice C. Stokes	3732					
The MAILING DATE of this communication appears on the cover sheet with the correspondence address							
Period for Reply A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS,							
WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION. - Extensions of time may be available under the provisions of 37 CFR 1.36(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication. - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).							
Status							
1) Responsive to communication(s) filed on 05 No.	ovember 2007.						
2a) This action is FINAL . 2b) ⊠ This	This action is FINAL . 2b)⊠ This action is non-final.						
	Since this application is in condition for allowance except for formal matters, prosecution as to the merits is						
closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213.							
Disposition of Claims							
4)⊠ Claim(s) <u>1-7,9-26,50 and 51</u> is/are pending in the application.							
4a) Of the above claim(s) is/are withdrawn from consideration.							
5) Claim(s) is/are allowed.							
	6)⊠ Claim(s) <u>1-7,9-26,50 and 51</u> is/are rejected.						
•	7) Claim(s) is/are objected to.						
8) Claim(s) are subject to restriction and/or election requirement.							
Application Papers							
9)☐ The specification is objected to by the Examiner.							
10) ☐ The drawing(s) filed on is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.							
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).							
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d). 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.							
The bath of declaration is objected to by the Examiner. Note the attached Onice Action of John F10-132.							
Priority under 35 U.S.C. § 119							
12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of:							
1. Certified copies of the priority documents have been received.							
2. Certified copies of the priority documents have been received in Application No							
3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).							
* See the attached detailed Office action for a list of the certified copies not received.							
·							
Attachment(c)							
Attachment(s) 1) Notice of References Cited (PTO-892) 4) Interview Summary (PTO-413)							
2) Notice of Draftsperson's Patent Drawing Review (PTO-948) Paper No(s)/Mail Date							
Information Disclosure Statement(s) (PTO/SB/08) Paper No(s)/Mail Date	5) Notice of Informal F 6) Other:	· · · · · · · · · · · · · · · · · · ·					

Art Unit: 3732

DETAILED ACTION

Continued Examination Under 37 CFR 1.114

A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on 11/05/07 has been entered.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(e) the invention was described in a patent granted on an application for patent by another filed in the United States before the invention thereof by the applicant for patent, or on an international application by another who has fulfilled the requirements of paragraphs (1), (2), and (4) of section 371(c) of this title before the invention thereof by the applicant for patent.

The changes made to 35 U.S.C. 102(e) by the American Inventors Protection Act of 1999 (AIPA) and the Intellectual Property and High Technology Technical Amendments Act of 2002 do not apply when the reference is a U.S. patent resulting directly or indirectly from an international application filed before November 29, 2000. Therefore, the prior art date of the reference is determined under 35 U.S.C. 102(e) prior to the amendment by the AIPA (pre-AIPA 35 U.S.C. 102(e)).

1) Claims 1-7,9-10,12-26 and 50-51 are rejected under 35 U.S.C. 102(e) as being anticipated by Morison et al. (USPN 6,296,642). Morrison et al. discloses a medical implant

09/644,777

Art Unit; 3732

with head 11, spaced arms and, a "closure member 12 adapted to be threadedly engaged" having a "generally cylindrical" portion, a clear implication that the threads are helically wound about the body of the implant in a continuous manner (column 3, lines 50 & 66). In column 3, lines 54-59, Morrison et al further disclose a second implant, "a receiver member 11" which "includes a longitudinal or thread axis". This device also includes "a transverse channel 16 for receiving an elongated member, which is generally perpendicular to axis 14 and bore 15". In lines 35-42 of column 4, Morrison et al disclose that "threads 44 and 46 are reverse angle threads" meaning that the "rearward-facing thread surface" is "sloped so that, for a given cross-section of the thread through the longitudinal axis of the screw, a point is closer to the distal or forward end of the screw that a point on the rearward-facing thread surface at the crest of the thread". Further, with respect to Claims 2-6, "in one particular embodiment of the present invention, illustrated in Fig. 4, pressure angle α is -5 degrees, and flank angle β is 45 degrees. However, it is understood that one of ordinary skill in the art would recognize that other negative values of pressure angle α , including values between about -1 degree and at least -40 degrees, and other values of flank angle β are within the scoped of the present invention. As noted above, reverse angel thread 44 of receiver member 11 is configured substantially similarly to reverse angle thread 46 of closure member 12 so that threads 44 and 46 can be engaged" (col. 4, lines 64-67 & col. 5, lines 1-7). In addition, "channel 16 is bounded on both sides by legs 20 of receiver member 11". In regards to Claims 14-16, as shown in Figures 1-3 of Morrison et al, the leading and trailing surfaces are spaced at substantially the same radius creating a generally obtuse triangular cross-section. Referring again to the Figures presented by Morrison et al, it is clear that Claims 21 are

Art Unit: 3732

anticipated by Figures 1-3 where β and α are shown to create a cross-section generally triangular in shape.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claim 11 is rejected under 35 U.S.C. 103(a) as being unpatentable over Morrison et al in view of Reed (USPN 5,499,892). Morrison et al discloses the claimed invention except for the threadform is a discontinuous helical pattern. Reed teaches a threadform with a helical discontinuous pattern. It would have been obvious to one having ordinary skill in the art at the time of the invention to incorporate the helical discontinuous pattern of the threadform taught by Reed into the threaded implant disclosed by Morrison et al in order to prevent easy removal of the closure from the implant thereby providing an increased interference fit.

Response to Amendment

The Declaration filed on 04/19/07 under 37 CFR 1.131 and the RCE filed 11/05/07 have been considered but both are ineffective to overcome the Morrison reference. The evidence submitted is insufficient to establish a reduction to practice of the invention in this country or a NAFTA or WTO member country prior to the effective date of the Morrison reference. The evidence submitted with the Declaration is insufficient to establish a reduction to practice prior to the effective filing date of the Morrison reference. Applicant's arguments submitted 11/05/07

have also been considered and are non-persuasive. The Declaration provides some evidence to support a conception of the invention prior to the effective filing date of the Morrison reference however, there is insufficient evidence provided that supports a reduction to practice prior to the effective filing date of the Morrison reference. As such, there is insufficient evidence to establish that the present invention was invented prior to the Morrison filing date.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Candice C. Stokes whose telephone number is (571) 272-4714. The examiner can normally be reached on 8:00am - 4:30pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Cris Rodriguez can be reached on (571) 272-4964. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Application/Control Number: 09/644,777

Art Unit: 3732

Candice C. Stokes

/Candice C. Stokes/, Examiner

Page 6

SUPERVISORY PATENT EXAMINER TECHNOLOGY CENTER 3700